

## Evaluating the AD9670 Octal Ultrasound AFE with Digital Demodulator

### FEATURES

Fully featured evaluation board for the [AD9670](#)  
PC software for controlling the [AD9670](#) via a USB interface

### EVALUATION KIT CONTENTS

**AD9670EBZ evaluation board**  
**1 switching power supply (6.0 V, 2.5 A), CUI EPS060250UH-PHP-SZ (provided)**

### EQUIPMENT NEEDED

Analog signal source and antialiasing filter  
Linear benchtop dc voltage source (0 V to 1.6 V), not required for CW Doppler mode  
PC running Microsoft Windows® 7 and Windows 10 operating system  
USB 2.0 port, recommended (USB 1.1 compatible)  
**HSC-ADC-EVALCZ FPGA-based data capture kit**  
1 switching power supply (6.0 V, 2.5 A), CUI EPS060250UH-PHP-SZ (provided with data capture kit)  
Spectrum analyzer for CW Doppler mode

### DOCUMENTS NEEDED

[AD9670](#) data sheet

### SOFTWARE NEEDED

[VisualAnalog](#)  
[SPIController](#)

### GENERAL DESCRIPTION

The AD9670EBZ evaluation board assists in evaluating all features of the [AD9670](#) octal ultrasound analog front end (AFE) device. The [AD9670](#) is low cost, low power, compact, and easy to use with medical ultrasound applications. The device provides eight voltage gain amplifier (VGA) channels, each with a low noise amplifier (LNA), a continuous wave (CW) harmonic rejection I/Q demodulator with programmable phase rotation, an antialiasing filter, an analog-to-digital converter (ADC), and a digital demodulator and decimator for both data processing and bandwidth reduction.

The evaluation board provides eight Subminiature Version A (SMA) connector inputs for all eight [AD9670](#) channels. The evaluation board mates with a field-programmable gate array (FPGA) data capture board ([HSC-ADC-EVALCZ](#)) that enables data capture through a USB connection to a computer.

The [SPIController](#) software provides flexible configuration of the [AD9670](#) and the [VisualAnalog](#)® software offers powerful data capture and signal analysis tools.

### EVALUATION BOARD PHOTOGRAPH




Figure 1.

## TABLE OF CONTENTS

|                                  |   |  |   |
|----------------------------------|---|--|---|
| Features .....                   | 1 | Evaluation Board Hardware.....               | 3 |
| Evaluation Kit Contents.....     | 1 | Power Supplies.....                          | 3 |
| Equipment Needed.....            | 1 | Input Signals.....                           | 3 |
| Documents Needed.....            | 1 | Output Signals .....                         | 3 |
| Software Needed .....            | 1 | Evaluation Board Software.....               | 4 |
| General Description .....        | 1 | Software Installation Procedures.....        | 4 |
| Evaluation Board Photograph..... | 1 | Evaluation Board Schematics and Artwork..... | 9 |
| Revision History .....           | 2 |  |   |

## REVISION HISTORY

1/2019—Revision 0: Initial Version

## EVALUATION BOARD HARDWARE

### POWER SUPPLIES

A switching power supply (6.0 V, 2.5 A) powers up the AD9670EBZ board.

### INPUT SIGNALS

The AD9670EBZ board provides eight SMA connector inputs for the eight AFE channels. The SMA channel inputs directly connect to the [AD9670](#) LNA input. An optional, external clock SMA input bypasses the default on-board 40 MHz crystal by changing the J301 jumper. Figure 1 shows an image of all inputs.

Other SMA inputs include the VGA  $V_{GAIN}$  control, the CW Doppler MLO $\pm$ , RESET $\pm$ , and an optional TX\_TRIG $\pm$  input. The  $V_{GAIN}$  single-ended signal is the default but a differential input configuration is also possible. The FPGA provides the TX\_TRIG $\pm$  signal on the [HSC-ADC-EVALCZ](#) data capture board.

### OUTPUT SIGNALS

The data capture board receives low voltage differential signaling (LVDS) digital output signals from the AD9670EBZ board through the on-board connectors (P401 and P402) (see Figure 1). The J501 jumper and J502 jumper SMA outputs are the CW Doppler summed in-phase quadrature (I/Q) outputs.

## EVALUATION BOARD SOFTWARE

### SOFTWARE INSTALLATION PROCEDURES

To install the VisualAnalog and SPIController software packages, and configure and test the AD9670, take the following steps:

1. Install the [VisualAnalog](#) software.
2. Install the [SPIController](#) software.
3. Plug the 6 V adapters into the AD9670EBZ evaluation board and the data capture board to power them up. Connect the data capture board to the PC with the USB cable that is provided with the data capture board.
4. Launch the [SPIController](#) software to open the window shown in Figure 2 and select the **AD9670.cfg** file from the dropdown list (under **File > Cfg Open Menu**). After the **Open File** dialog appears, browse for and select the **AD9670\_14bit\_125MSspiR03.cfg** file. The chip ID reads **AD9670** (see Figure 2) after loading the **AD9670.cfg** file.

**CHIP READS AD9670**





Figure 2. SPIController Main Window

5. From the **Config** menu, click **Launch Memory Map Dialog** to open the **Memory Map** window, as shown in Figure 3. Click the **Load File** button to open the **AD9670\_14bit\_125MSspiR03.mmp** file, which loads the coefficient and profile memory data into the **AD9670**.

READS THE MEMORY ON THE DEVICE AND SAVES IT TO A FILE  
DEMODULATION FREQUENCY CAN BE CHANGED BY UPDATING DEMOD AND SAMPLE FREQUENCY



17260-003

Figure 3. Memory Map Window

6. In the main [SPIController](#) window, open the **File** menu and click **Open Macrogroup**. Select the **AD9670\_Initialization.mgp** macro file to open the **MacroEditor** window shown in Figure 4. The macro file is an .XML file that lists the serial peripheral interface (SPI) command macros in the **D6970\_Initialization\_v2.mg** dropdown box (see Figure 4). Each command macro includes SPI writes that configure the **AD9670** for specific operating modes. The **MacroEditor** conducts a batch write of SPI registers to set up the **AD9670**. The macro file contains the following two macros: **Real Mode, 40Msps** and **IQ Mode, 40Msps**. It is possible to add unique configuration macros to the .XML macro file and load the macros into the **MacroEditor** window.




Figure 4. MacroEditor Window and D6970\_Initialization\_v2.mg Dropdown Box




Figure 5. Select, Enable, and Run Commands in MacroEditor

17280-005

Table 1. Procedural List of SPI Commands Configuring the AD9670

| Register     | Function   |
|--------------|--|
| 0x000        | SPI reset.   |
| 0x002        | Set speed mode.  |
| 0x004, 0x005 | Set local registers to all channels.   |
| 0x011        | Set LNA, VGA, and PGA gain.  |
| 0x10C        | Set SPI TX_TRIGGER± and Index Profile 1.                                       |
| 0x113        | Enable/bypass digital filters, demodulator, and decimator.                     |
| 0x00F        | Set low-pass filter cutoff frequency and mode.                                 |
| 0x021        | 16-bit, 8-lane, frame clock output (FCO) covers entire frame.                  |
| 0x008        | Time gain compensation (TGC) run mode.   |
| 0x182        | Autoconfigure phase-locked loop (PLL).   |
| 0x199        | Autoset FCO and data clock output (DCO).                                       |
| 0x02B        | Set analog low-pass filter and high-pass filter to defaults and tunes filters. |
| 0x014        | Set output data format.  |
| 0x188        | Enable start code.   |
| 0x10C        | Set SPI TX_TRIGGER± and Index Profile 1.                                       |
| 0x19B        | Serial format.   |

- Select either the **Real Mode** or the **IQ Mode** macro from the dropdown box (see Figure 5) and select the **Enable** checkbox. To run the selected macro, click the **Run** button to conduct a write operation (see Figure 5).

- Launch the [VisualAnalog](#) software to use the canvas templates listed in the [AD9670](#) directory in the **Categories** pane (see Figure 6). The **Samples** canvas shows time domain output samples vs. sample count.




Figure 6. VisualAnalog Canvas Selection Window

17280-006

9. If the **VisualAnalog** software opens and only shows the task bar, as shown in Figure 7, click the down arrow on the right side of the window to expand the **AD9670 Samples** canvas (shown in Figure 8) to see the signal processing flow.




Figure 7. **VisualAnalog** in Window Collapse Mode

10. Next, load the FPGA binary file into the hardware. To load the file, click the **Settings** button in the **ADC Data Capture** block (see Figure 8).




Figure 8. **VisualAnalog** Sample Canvas

11. When the ADC Data Capture Settings window opens, click the **Capture Board** tab and then click the **Browse...** button. Select the **ad9670\_lvds\_mar2012.bin** file and click **Program**, which illuminates the **Done** LED D6 indicator on the **HSC-ADC-EVALCZ** board. Click **OK** to close the window.




Figure 9. ADC Data Capture Settings Window

12. The **AD9670** is now running and fully configured for data capture mode. The **VisualAnalog** software offers two data capture modes: onetime and infinite loop. Click the blue right arrow (see Figure 7 or Figure 8) to run the canvas in onetime mode and click the green repeat button (see Figure 7 or Figure 8) to run the canvas in infinite loop mode. The **Samples** canvas then opens two windows showing 8192 time domain samples for Channel A and Channel B (see Figure 10 for the Channel A window).

#### BUTTONS FOR ZOOMING IN AND RESIZING




Figure 10. Data Capture Window

17280-010

13. Another loadable canvas, the **AD9670\_Average\_FFT.vac** canvas, runs average fast Fourier transform (FFT) calculations of the captured samples. The average count is set to five captures. Therefore, either run the canvas five times to display the first FFT output or leave it in continuous run mode to keep the canvas functioning.
14. Figure 11 shows both the FFT output and how the analog and digital high-pass filters shape the noise floor in the low frequency region.




Figure 11. Average FFT Canvas

17280-011

15. To disable the digital high-pass filter, select the **HP Filter Bypass** checkbox in the **Memory Map** window (see Figure 12) and then click the **Manual Tx Trigger** button (see Figure 12).




Figure 12. Disabling the Digital High-Pass Filter Bypass in the **Memory Map** Window




Figure 13. Average FFT Canvas with the Digital High-Pass Filter Disabled

17. To enable the demodulator and decimator, select and run the **IQ Mode, 40Msps** macro in the **MacroEditor** window (see Figure 4).

## EVALUATION BOARD SCHEMATICS AND ARTWORK

17260-015




Figure 14. AD9670EBZ Schematics Page 1




Figure 15. AD9670EBZ Schematics Page 2





Figure 16. AD9670EBZ Schematics Page 3






Figure 18. AD9670EBZ Schematics Page 5

Rev. 0 | Page 13 of 15




Figure 19. AD9670EBZ Schematics Page 6

## NOTES



### ESD Caution

**ESD (electrostatic discharge) sensitive device.** Charged devices and circuit boards can discharge without detection. Although this product features patented or proprietary protection circuitry, damage may occur on devices subjected to high energy ESD. Therefore, proper ESD precautions should be taken to avoid performance degradation or loss of functionality.

### Legal Terms and Conditions

By using the evaluation board discussed herein (together with any tools, components documentation or support materials, the "Evaluation Board"), you are agreeing to be bound by the terms and conditions set forth below ("Agreement") unless you have purchased the Evaluation Board, in which case the Analog Devices Standard Terms and Conditions of Sale shall govern. Do not use the Evaluation Board until you have read and agreed to the Agreement. Your use of the Evaluation Board shall signify your acceptance of the Agreement. This Agreement is made by and between you ("Customer") and Analog Devices, Inc. ("ADI"), with its principal place of business at One Technology Way, Norwood, MA 02062, USA. Subject to the terms and conditions of the Agreement, ADI hereby grants to Customer a free, limited, personal, temporary, non-exclusive, non-sublicensable, non-transferable license to use the Evaluation Board FOR EVALUATION PURPOSES ONLY. Customer understands and agrees that the Evaluation Board is provided for the sole and exclusive purpose referenced above, and agrees not to use the Evaluation Board for any other purpose. Furthermore, the license granted is expressly made subject to the following additional limitations: Customer shall not (i) rent, lease, display, sell, transfer, assign, sublicense, or distribute the Evaluation Board; and (ii) permit any Third Party to access the Evaluation Board. As used herein, the term "Third Party" includes any entity other than ADI, Customer, their employees, affiliates and in-house consultants. The Evaluation Board is NOT sold to Customer; all rights not expressly granted herein, including ownership of the Evaluation Board, are reserved by ADI. CONFIDENTIALITY. This Agreement and the Evaluation Board shall all be considered the confidential and proprietary information of ADI. Customer may not disclose or transfer any portion of the Evaluation Board to any other party for any reason. Upon discontinuation of use of the Evaluation Board or termination of this Agreement, Customer agrees to promptly return the Evaluation Board to ADI. ADDITIONAL RESTRICTIONS. Customer may not disassemble, decompile or reverse engineer chips on the Evaluation Board. Customer shall inform ADI of any occurred damages or any modifications or alterations it makes to the Evaluation Board, including but not limited to soldering or any other activity that affects the material content of the Evaluation Board. Modifications to the Evaluation Board must comply with applicable law, including but not limited to the RoHS Directive. TERMINATION. ADI may terminate this Agreement at any time upon giving written notice to Customer. Customer agrees to return to ADI the Evaluation Board at that time. LIMITATION OF LIABILITY. THE EVALUATION BOARD PROVIDED HEREUNDER IS PROVIDED "AS IS" AND ADI MAKES NO WARRANTIES OR REPRESENTATIONS OF ANY KIND WITH RESPECT TO IT. ADI SPECIFICALLY DISCLAIMS ANY REPRESENTATIONS, ENDORSEMENTS, GUARANTEES, OR WARRANTIES, EXPRESS OR IMPLIED, RELATED TO THE EVALUATION BOARD INCLUDING, BUT NOT LIMITED TO, THE IMPLIED WARRANTY OF MERCHANTABILITY, TITLE, FITNESS FOR A PARTICULAR PURPOSE OR NONINFRINGEMENT OF INTELLECTUAL PROPERTY RIGHTS. IN NO EVENT WILL ADI AND ITS LICENSORS BE LIABLE FOR ANY INCIDENTAL, SPECIAL, INDIRECT, OR CONSEQUENTIAL DAMAGES RESULTING FROM CUSTOMER'S POSSESSION OR USE OF THE EVALUATION BOARD, INCLUDING BUT NOT LIMITED TO LOST PROFITS, DELAY COSTS, LABOR COSTS OR LOSS OF GOODWILL. ADI'S TOTAL LIABILITY FROM ANY AND ALL CAUSES SHALL BE LIMITED TO THE AMOUNT OF ONE HUNDRED US DOLLARS (\$100.00). EXPORT. Customer agrees that it will not directly or indirectly export the Evaluation Board to another country, and that it will comply with all applicable United States federal laws and regulations relating to exports. GOVERNING LAW. This Agreement shall be governed by and construed in accordance with the substantive laws of the Commonwealth of Massachusetts (excluding conflict of law rules). Any legal action regarding this Agreement will be heard in the state or federal courts having jurisdiction in Suffolk County, Massachusetts, and Customer hereby submits to the personal jurisdiction and venue of such courts. The United Nations Convention on Contracts for the International Sale of Goods shall not apply to this Agreement and is expressly disclaimed.