
THESE DRAWINGS ARE THE PROPERTY OF QUALTEK ELECTRONICS AND NOT TO BE REPRODUCED FOR OR COMMUNICATED TO A THIRD PARTY WITHOUT THE EXPRESS WRITTEN PERMISSION OF AN OFFICER OF QUALTEK ELECTRONICS.

REVISION					
REV.	DESCRIPTION	DATE	APPROVED		
В	RoHS Compliancy	06-01-06	DP		

## $120'' \pm \frac{2''}{0''}$



## CONTROLLED

## NOTES:

1.	CORDAGE: SJT, 18AWG / 3C , UNSHIELDED,
	NA COLOR CODE, TEMP. RATING 60°C,
	RATING: 125V, 10A, JACKET COLOR: GRAY
2.	PLUG: NEMA 5-15P HOSPITAL GRADE (CLEAR)
3.	CONNECTOR: IEC 60320 C13
4.	APPROVALS: UL, CSA
5.	ROHS COMPLIANT

6. CAUTION LABEL

Dim range	Tolerance +/-	Qualtek Electronics Corp. PPC DIVISION			
(mm)	(mm)	Part Number:	233003-06	RoHS Compliant	
0-1	0.15	Description: POWER CORD HOSPITAL GRADE			
1-6	0.20				
6-18	0.30				
18-50	0.50		UNIT: in	REV. B	
50-120	0.70	Drawn / Date	Checked / Date	Approved / Date	
120-250	0.90	B.Zhang	10-21-03	JJH / 10-21-03	
250-500	1.20	10-21-03		5517 10-21-05	

## **Mouser Electronics**

Authorized Distributor

Click to View Pricing, Inventory, Delivery & Lifecycle Information:

Qualtek: 233003-06